



Material and Engineering Laboratory-Taipei

TEST REPORT



REPORT NO. : HV-20-07376A

PAGE : 2 OF 3

REPORT DATE : DEC. 30, 2020

Direct Screw Withdrawal

Test Equipment :

Name	Brand	Model
50kN Universal Testing Machine	CHUN YEN	-

Test Method :

Ref. ASTM D1037-12(2020) Standard Test Methods for Evaluating Properties of Wood-Base Fiber and Particle Panel Materials

Test Condition :

Test Speed : 1.5 mm/min

Depth of screw penetration : 25.6 mm

Test Result :

Test Item(s)	Test Result(s)
Direct Screw Withdrawal (kgf)	329

----- 2 -----

The value of required specifications are for reference only.

Conformity judgement is the Applicant's final verdict.

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com.tw/Terms-and-Conditions> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com.tw/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested.